

CONFIDENTIAL FINAL REPORT**SPONSOR:** GWELLKOREA.CO.LTD**SPONSOR'S REPRESENTATIVE:** Warren Kim**STUDY TITLE:** EVALUATION OF INACTIVATION OF AIRBORNE VIRUS BY AN AIR DISINFECTION DEVICE – Human Coronavirus (Similar to MERS-Coronavirus and SARS-Coronavirus)**STUDY IDENTIFICATION:** MicroBioTest Project No. 916-101 (refer to signed protocol)

TEST AGENT NAME	LOT NO.	DATE RECEIVED	DS NO.
UPI Device (LUVA Air Purifier)	N/A	03/09/16	G122

ACTIVE INGREDIENT(S): Positive/negative ions**CHALLENGE ORGANISM:** Human Coronavirus, strain 229E; ATCC VR-740**HOST CELL LINE:** MRC-5 cells, ATCC CCL-171**DILUTION MEDIUM:** 1X Minimum Essential Medium (MEM) + 2% Fetal Bovine Serum (FBS)**SEMI-SOLID COLLECTION MEDIUM :** 1X MEM + 5% Gelatin + 1% FBS + 1% HEPES + 10 µg/mL Gentamicin + 1% NaHCO₃ + 1% Penicillin-Streptomycin + 2.5 µg/mL Amphotericin B**FLUSH MEDIUM :** 1X MEM + 1% FBS + 1% HEPES + 10 µg/mL Gentamicin + 1% NaHCO₃ + 1% Penicillin-Streptomycin + 2.5 µg/mL Amphotericin B**AEROSOL MEDIUM:** 0.1X MEM**VIRUS AEROSOL TIME:** 25 minutes**EXPOSURE (CONTACT) TIME:** 10 minutes and 30 minutes

EXPOSURE TEMPERATURE:	Ambient temperature
NUMBER OF REPLICATES:	Two replicates per contact time at four wells per dilution
INCUBATION TEMPERATURE:	33±2°C with 5±1% CO ₂
INCUBATION TIME:	5 – 7 days (7 days actual)
DEVICE CONNECTION:	The UPI Device (LUVA Air Purifier) was connected to the adapter and plugged into a power supply to operate.

CALCULATION OF TITER:

The 50% tissue culture infectious dose per mL (TCID₅₀/mL) was determined using the Spearman-Kärber method using the following formula:

$$m = x_k + \left(\frac{d}{2} \right) - d \sum p_i$$

where:

- m = the logarithm of the dilution at which half of the wells are infected relative to the test volume
- x_k = the logarithm of the smallest dosage which induces infection in all cultures
- d = the logarithm of the dilution factor
- p_i = the proportion of positive results at dilution i
- Σp_i = the sum of p_i (starting with the highest dilution producing 100% infection)

The values were converted to TCID₅₀/mL using a sample inoculum of 1.0 mL.

RESULTS:

Results are presented in Tables 1-2.

The Viral load was determined in the following manner:

$$\text{Viral Load (Log}_{10} \text{ TCID}_{50}) = \text{Titer (Log}_{10} \text{ TCID}_{50}/\text{mL}) + \text{Log}_{10}[\text{Volume (mL)}]$$

The log₁₀ Reduction Factor (LRF) was calculated in the following manner:

$$\text{Log}_{10} \text{ Reduction Factor} = \text{Initial viral load (Log}_{10} \text{ TCID}_{50}) - \text{Output viral load (Log}_{10} \text{ TCID}_{50})$$

RESULTS (Continued):

The Mean Viral Log₁₀ Reduction from n replicates was determined as follows:

$$\text{Log}_{10} \left[\frac{10^{(\text{Log}_{10}\text{Load}_1)} + 10^{(\text{Log}_{10}\text{Load}_2)} + \dots + 10^{(\text{Log}_{10}\text{Load}_n)}}{n} \right]$$

The Percentage of Virus Inactivation was calculated in the following manner:

$$[1 - \text{Output Viral Load} / \text{Initial Viral Load}] \times 100 = 1 - 10^{(-\text{Log}_{10} \text{Reduction Factor})} \times 100$$

Table 1
Titer Results

Sample	Atomizer time	Replicates	Titer (Log ₁₀ TCID ₅₀ /mL)	Volume (mL)	Viral Load (Log ₁₀ TCID ₅₀)
Cell viability/media sterility control	N/A	N/A	no virus detected, cells viable; media sterile		
Volume application evaluation			average volume of challenge per run: 10.1 mL		
Virus Stock Titer Control			7.00	-	-
Virus Input Load Control	N/A	1	4.75	10	5.75
		2	5.00	10	6.00
	Average				5.89
UPI Device (LUVA Air Purifier)*	10 minutes	1	2.75	10	3.75
		2	3.00	10	4.00
	30 minutes	1	1.50	10	2.50
		2	1.75	10	2.75

*Cytotoxicity observed at the undiluted dilution

NA = Not Applicable

RESULTS (Continued):

Table 2
Viral Reduction

Test Agent(s)	Atomizer Time	Replicate Number	Initial Viral Load* (Log ₁₀ TCID ₅₀)	Output Viral Load (Log ₁₀ TCID ₅₀)	Log ₁₀ Reduction	Percentage Log ₁₀ Reduction
UPI Device (LUVA Air Purifier)	10 minutes	1	5.89	3.75	2.14	99.1
		2		4.00	1.89	
	Mean Reduction ± 95% Confidence Interval				2.04 ± 0.25	
	30 minutes	1	5.89	2.50	3.39	99.9
		2		2.75	3.14	
	Mean Reduction ± 95% Confidence Interval				3.29 ± 0.25	

* Results represent the average of two replicates.

CONCLUSION:

GWELLKOREA.CO.LTD UPI Device (LUVA Air Purifier) was evaluated for its ability to inactivate Human Coronavirus (Similar to MERS-Coronavirus and SARS-Coronavirus). MicroBioTest personnel performed the inactivation procedure using Human Coronavirus, Strain: 229E. Samples were titrated by 50% tissue culture infectious dose (TCID₅₀) endpoint assay using MRC-5 cells.

The viral reductions for the GWELLKOREA.CO.LTD UPI Device (LUVA Air Purifier) are presented in Table 2. All of the controls met the criteria for a valid test.

Study Director: _____

S. Steve Zhou, Ph.D.

06/08/2016

Date